



# Ultraviolet 3.2g Nitrile Powder Free Examination Glove

# Nitrile Powder Free Examination Glove (ASTM & EN)

Product S	oecification	
Product Code	N032ULV-CS-24FT	
Product Type	Non-Sterile, Powder-Free, Nitrile, Ambidextrous, Examination Glove	
Color	Ultraviolet	
Intended Use	Medical use - for medical examinations and procedures to help prevent contamination between healthcare workers and patients.	
Product Conformance This product meets the current version of ASTM D6319, EN 44 and Customer Requirements / Category I - PPE (EU 2016/42E EN ISO 21420:2020 / Food Contact: UNI EN1186, EN13130. This product is also manufactured in accordance with a qualisystem that conforms to ISO 9001 and ISO 13485.		
Material	Nitrile Butadiene Rubber	
Cuff Finishing	Beaded	
Surface Finishing	Finger Textured	
Other Requirements	Other requirements, including labelling and packaging, is subject to customer and market regulatory requirements and is specified in the Customer Product Sheet.	
Certificate of Release	A Certificate of Release is available upon request to ensure all products confirm to specification.	

Requirement: PAH content per EN ISO 21420:2020	< 0.2 mg/kg (of each PAH listed in the appendices)	Pass
Requirement: REACH 1907/2006 annex XVII entry number 50	< 0.01% (w/w) of each SVHC Candidate List	Pass

# Physical Dimension

Size	XS	S	М	L	XL
Palm Width (mm)	75 ± 5	85 ± 5	95 ± 5	105 ± 5	115 ± 5
Length (mm) (min.)	240 mm				
Weight (g)	3.2 ± 0.3 g				
Location of Thickness Measurement	Single Wall (mm)				
~ Cuff (25 ± 5 mm from cuff edge)	≥ 0.04				
~ Palm (33 ± 5 mm from crotch)	≥ 0.05				
~ Finger (13 ± 3 mm from fingertip)	≥ 0.06				

## Physical Properties

Criteria	Before Aging	After Aging
Tensile Strength (Mpa) (min.)	14	14
Elongation (%) (min.)	500	400
Force at Break (N) (Median)	6	6

## Pre-shipment Inspection

Characteristic	Dimension	Physical Properties	Freedom from Holes	Visual Defects	
				Major	Minor
Inspection Level	\$2	S2	G1	G1	G1
AQL	4.0	4.0	1.5	2.5	4.0











